

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 16 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130305 ✓
A	Name	Anthony Wallace	ID# 6555 Date 6/10/10 ✓
	Agency	Hoona Police Dept	Phone # 945-3655
	Instrument Location	Hoona Police Dept ✓	
B	Alco S/N	X 301532	Target Value .082 ✓ High Pressure
	Alco Test Value Average	.089 ✓	.088 ✓
		1 st Alco	2 nd Alco
Signature		[Signature] ✓	
(OVER)			

CAB
6/18/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Do Not fill in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of July, 2010.

[Signature]
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130305 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305 ✓

JUNE 10, 2010 ✓

OPERATOR'S NAME:

WALLACE ANTHONY M

OPERATOR'S NUMBER: 6555

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: HHM1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X301532 ✓

— BREATH ANALYSIS —

.002 ADJUSTED FOR 30.17 in		
ALCO TARGET	.002	10:34
BLANK TEST	.000	10:35
INTERNAL STANDARD	VERIFIED	10:35
ALCO TV 30.17 in	.009 ✓	10:35
BLANK TEST	.000	10:36
SUBJECT SAMPLE	.000	10:36
BLANK TEST	.000	10:37
ALCO TV 30.19 in	.008 ✓	10:37
BLANK TEST	.000	10:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305 ✓

JUNE 10, 2010 ✓

TIME 10:30 ✓

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	30.17 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!\"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJ KLMNOPQRSTUVWXYZ[\]^_`abc defghijklmno
pqrstuvwxyz{|}~||